

Quality certificate  
Technical data for Validation  
of LCK333 (0.2-6.0 mg/l  
nonionic. Surfactants)

Quality certificate

Technical data for cuvette test LCK333  
(Results as Triton X-100)

Sensitivity	0.146 Abs./(mg/l)
Ordinate intersect	0.087 Abs.
Residual standard deviation	0.0146 Abs.
Method variation coefficient	3.05 %
Method standard deviation	0.10 mg/l
Confidence intervall (95%)	$\pm 0.243$ mg/l

The technical data for cuvette test LCK333 were determined in conformity with ISO 8466-1 and DIN 38402 A51 „Calibration of analysis methods“.


The series of the smallest and largest calibration standards exhibit normal distribution and are outlier- and trend-free.

The calibration gives a linear function.

Technical data in conformity with DIN 32645

Detection limit	0.02 mg/l
Quantitation limit	0.06 mg/l

The detection and the quantitation limits were determined in conformity with DIN 32645.



Result	Confidence intervall (95%)
1.20 mg/l	$\pm 0.259$ mg/l
2.40 mg/l	$\pm 0.246$ mg/l
3.60 mg/l	$\pm 0.244$ mg/l
4.80 mg/l	$\pm 0.251$ mg/l
6.00 mg/l	$\pm 0.267$ mg/l

**HACH LANGE GmbH**  
Quality Management

**Dr. Ralf Kloos**