



Reference : 01-029
Product :
KING B AGAR (F AGAR)

Scharlau Microbiology - Technical data sheet

Also known as

Pigment Production Agar B; Ps Medium B; Fluorescein Agar; F Agar; Flo Agar; Pseudomonas Agar Medium for Detection of Fluorescein

Specification

Culture media for enhancing the fluorescein production by *Pseudomonas spp.* according to ISO standards.

Formula * in g/L

Meat peptone..... 10.0
Casein peptone..... 10.0
Dipotassium phosphate..... 1.5
Magnesium sulfate..... 1.5
Agar..... 15.0

Final pH 7,2 ±0,2 at 25 °C

* Adjusted and /or supplemented as required to meet performance criteria

Directions

Suspend 38 g of powder in 1 L of distilled water with 10 mL of glycerol and let it soak. Heat to boiling and distribute in suitable containers. Sterilize in the autoclave at 121°C for 15 minutes. Cool by solidifying in slanted position with a long slant.

Description

F Medium was formulated by King, Ward and Raney in 1954 to enhance green fluorescent pigment (pyoverdine) production by *Pseudomonas fluorescens* and *P. aeruginosa*, in which pyocyanin production is restricted.

Green-yellowish pigments, soluble and fluorescent, define *Pseudomonas* group I according to the 9th edition of Bergey's Manual of Systematic Bacteriology, and therefore, detection of their production is critical.

Technique

Slanted tubes are inoculated with *Pseudomonas* strains and incubated à 30-32°C for a 2-4 days period. If after this time a green-yellowish colour does not appear on the medium, the tubes should be kept under observation à room temperature for an additional period of 6-20 days before the culture can be regarded as negative. It should be noted that *Pseudomonas aeruginosa* and *Pseudomonas putida* strains obtained from water, soil or food, produce pigments slowly.

Pyoverdine is not soluble in chloroform, so the confirmation of its presence is usually done by a characteristic fluorescence verification under Wood's light (365 µm), comparing the suspected positive tube to another un-inoculated F Medium tube, which is considered as the control.

Quality control

Incubation temperature: 30 ± 1°C

Incubation time: 44 ± 4h

Inoculum: Pure culture is inoculated by surface streaking, according to ISO 11133:2014/Amd 1:2018 & Adm 2:2020

Microorganism

Growth

Remarks

<i>Pseudomonas fluorescens</i> ATCC® 13525	Good to very good	F (+)
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Good to very good	Yellow-green
<i>Pseudomonas aeruginosa</i> ATCC® 9027	Good to very good	Yellow-green
<i>Pseudomonas aeruginosa</i> ATCC® 10145	Good to very good	Yellow-green
<i>Burkholderia cepacia</i> ATCC® 25608	Good to very good	Without pigment
<i>Escherichia coli</i> ATCC® 8739	Good	F (-)



Reference : 01-029

Scharlau Microbiology - Technical data sheet

Product :
KING B AGAR (F AGAR)

References

- DIN 38411 Standard (1991) Parte 6: Mikrobiologischen Verfahren (Gruppe K) Nachweis von Escherichia coli und coliformen keimen (K6).
- ISO 16266 Standard (2006) Water Quality. Detection and enumeration of *Ps aeruginosa*. Method by membrane filtration.
- ISO 11133:2014/ Adm 1:2018/ Adm 2:2020/ Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO 22717 Standard (2015) Cosmetics - Microbiology - Detection of *Pseudomonas aeruginosa*.
- KING, E.O., M.WARD & D.E. RANEY (1954) Two simple media for the demonstration of pyocyanin and fluorescein J. Lab.Clin.Med. 44:30-307.
- LENNETTE, E.H., E.W. SPAULDING & J.P. TROUANT (1974) Manual of Clinical Microbiology. 2nd ed. ASM. Washington.
- PALLERONI, N. (1984) The genus *Pseudomonas*, in Bergey's Manual of Systematic Bacteriology.
- USP (2008) 31th ed. <61> Microbial Limit Tests. US Pharmacopeial Convention Inc. Rockville. MD.

Storage

For laboratory use only. Keep tightly closed, away from bright light, in a cool dry place (+4 °C to 30 °C).
