



Reference : 02-725

Scharlau Microbiology - Technical data sheet

Product :
TSB Non Animal - Irradiated

Also known as

Vegetal TSB

Specification

Highly nutritious general purpose liquid medium, formulated according to harmonized pharmacopoeia, where none of the components is of animal origin.

Formula * in g/L

Peptones 20.00
Sodium chloride..... 5.00
Dipotassium phosphate.....2.50
Dextrose..... 2.50

Final pH 7,3±0,2 at 25 °C

* Adjusted and /or supplemented as required to meet performance criteria

Directions

For liquid filling tests: Suspend 30 g in 1 L of sterile distilled water and mix well to dissolve. Use for liquid fill validation procedures.

For solid filling tests: Use powder for dry-fill validation procedures. Ensure that the final concentration of the medium is 30 g of Irradiated TSB suspended in 1 L of sterile distilled water.

Description

A general purpose culture medium formulated according to pharmacopoeias, which has replaced the peptone of animal origin (casein peptone), using plant and not animal sources. Obtaining the same end product and completely eliminating the potential risks associated with animal by-products of type C, such as the casein peptone.

The Irradiated TSB-Non Animal is the classical TSB culture medium sterilized by a gamma-irradiation process making it appropriate for validation of aseptic filling processes in the pharmaceutical industry (Media-Fill Test).

The sterility of the dehydrated culture medium is verified according to the methodology as described in the pharmacopoeia. It is subjected to exactly the same conditions as the pharmaceutical product, including filling and closing to ensure that there is no microbial contamination occurring during the process.

The gamma-irradiation treatment is carried out according to the Annex B of the ISO 11137 Standard, and warrants a 25 kGy absorption by the medium, which is enough to suppress any vegetative and/or spore form of microorganisms present in the powder without modification of the medium's performance. All batches of Irradiated TSB are checked for sterility and performance.

Technique

All the conditions and data for the validation of aseptic filling process can be consulted in the ISO Standard 13408-1:1998 in the chapters devoted to the methods of preparation of sterile products in several Pharmacopoeias.

Quality control

Incubation temperature: 30-35°C

Incubation time: 18-72h / ≤ 5 d

Inoculum: Practical range 10-100 CFU. Min. 50 CFU (Productivity) according to Eur. Pharm. harm.

Microorganism	Growth	Remarks
<i>Bacillus subtilis</i> ATCC® 6633	Good	≤ 3 days
<i>Staphylococcus aureus</i> ATCC® 6538	Good	≤ 3 days
<i>Escherichia coli</i> ATCC® 8739	Good	≤ 3 days
<i>Pseudomonas aeruginosa</i> ATCC® 9027	Good	≤ 3 days
<i>Candida albicans</i> ATCC® 10231	Good	≤ 5 days
<i>Salmonella abony</i> NCTC® 6017	Good	≤ 3 days
<i>Aspergillus brasiliensis</i> ATCC® 16404	Good	≤ 5 days



Reference : 02-725

Scharlau Microbiology - Technical data sheet

Product :

TSB Non Animal - Irradiated

References

- EUROPEAN PHARMACOPOEIA 10.0 (2020) 10th ed. § 2.6.13. Microbiological examination of non-sterile products: Test for specified microorganisms. Harmonised Method. EDQM. Council of Europe. Strasbourg.
- ISO Standard 11137: 1995. Sterilization of health care products - Requirements for validation and routine control - Radiation Sterilization. (Annex B).
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO Standard 13408-1: 1998. Aseptic processing of health care products - Part 1: General requirements.
- US PHARMACOPOEIA 28 / NATIONAL FORMULARY 23 (2005) General Chapters § <71> Sterility Tests y § <1208> Sterilization and Sterility Assurance of Compendial Articles.

Storage

For laboratory use only. Keep tightly closed, away from bright light, in a cool dry place (+4 °C to 30 °C).
