



Reference: 06-089-100

Scharlau Microbiology - Technical Data

Product: Potassium Tellurite Solution 1%

Specification

Aqueous solution of potassium tellurite at 1%, sterilized by filtration and suitable for use as an inhibitor additive in culture media.

Presentation

1 Bottle
125 ml bottle
with: 100 ± 3 ml

Packaging Details

1 box with 1 bottle (amber) 125 ml. Injectable cap:
Plastic screw inner cap. The use of syringes needles
with a diameter greater than 0.8 mm is not
recommended.

Shelf Life

18 months

Storage

15-25 °C

Composition

Potassium tellurite..... 10.0
Sterile water..... 1000 ml

Reagent to be added, as inhibitor, into
culture media for staphylococci isolation.

Description /Technique

Potassium Tellurite Solution is added to culture media as an inhibitor. Its purpose is to prevent the growth of most Gram negative bacteria and of those Gram positive bacteria unable to reduce it.

It is used in media such as Giolitti-Cantoni Broth, Vogel-Johnson Agar, Baird Parker and other selective media for staphylococci. This solution is also contained in selective media for corynebacteria, streptococci and vibrios.

There is a high correlation between the ability to reduce potassium tellurite to tellurium and the pathogenicity of staphylococci, therefore, the presence of potassium tellurite in a medium, together with other tests, helps to determine staphylococci of clinical interest. Potassium Tellurite Solution should be stored at room temperature, since low temperatures will cause crystallization and later precipitation of the product. Should this occur, intense agitation will help redissolve the sediment. Due to its thermolability, the potassium tellurite It is recommended not to subject it to temperatures above 40°C.

Note: This product is transparent but precipitates may appear without affecting performance of the medium

Quality control

Physical/Chemical control

Color : Transparent/colourless

Microbiological control

Add supplement to functionality - into Giolitti Cantoni Broth base

Inoculate: Practical range 100 ± 20 CFU. min. 50 CFU (productivity)/ 10⁴-10⁶ (selectivity).

Analytical methodology according to ISO 11133:2014/A1:2018; A2:2020.

Aerobiosis. Incubation at 37 °C ± 1, reading after 24-48 ± 2h

Microorganism

Stph. aureus ATCC® 25923, WDCM 00034

Staphylococcus aureus ATCC® 6538, WDCM 00032

Escherichia coli ATCC® 8739, WDCM 00012

Growth

Good - Black precipitate

Good - Black precipitate

Inhibited

Sterility control

Incubation 48 h at 30-35 °C and 48 h at 20-25 °C: NO GROWTH.

Check at 7 days after incubation in same conditions.



Reference: 06-089-100

Scharlau Microbiology - Technical Data

Product: Potassium Tellurite Solution 1%

Bibliography

- ATLAS R.M. & L.C. PARKS (1993) Handbook of Microbiological Media. CRC Press. London.
- BAIRD-PARKER, A.C. (1962) An improved diagnostic and selective medium for isolating coagulase-positive staphylococci. J. Appl. Bact. 25:12.
- COLIPA (1997) Guidelines on Microbial Quality Management (MQM). Brussels.
- DOWNES, F.P. & K. ITO (2001) Compendium of Methods for the Microbiological Examination of Foods. 4th ed. APHA. Washington. USA.
- EUROPEAN PHARMACOPOEIA (2007) 5thed. Suppl. 5.6 § 2.6.13 Microbiological examination of non-sterile products. EDQM. Council of Europe. Strasbourg.
- FIL-IDF 60:2001 Standard. Lait et produits à base de lait - Detection des staphylocoques à coagulase positive - Technique du nombre le plus probable. Brussels.
- ISO 5944:2001 Standard. Milk and Milk based products - Detection of coagulase positive staphylococci - MPN Technique. Geneva.
- ISO 6888-1:1999 Standard. Microbiology of food and animal feeding stuffs - Horizontal method for the enumeration of coagulase-positive staphylococci - Part 1 Technique using Baird-Parker Agar medium. Geneva.
- ISO 11133:2014/ Adm 1:2018. Microbiology of food, animal feed and water. Preparation, production, storage and performance testing of culture media.
- ISO 22718 Standard (2015) . Cosmetics - Microbiology - Detection of *Staphylococcus aureus*.
- USP 31 - NF 26 (2008) <61> Microbial Limit Tests. US Pharmacopoeial Conv. Inc. Rockville. MD. USA.
- ZANGERL, P. & H. ASPERGER (2003) Media used in the detection and enumeration of *Staphylococcus aureus*. In Handbook.